EC DECLARATION OF CONFORMITY
According to Annex V and VII of MDD 93/42/EEC

TF86 01 27 January 2015
GMDN 16704

GC EUROPE N.V.
Research Park
Interleuvenlaan 33
B-3001 Leuven
Belgium

We ensure and declare under our sole responsibility that the product :

GC FujiCEM 2

to which this declaration relates is in conformity with the following standards or other normative documents :

ISO 9917-2: 2010 Dentistry water-based cements – Part 2: Resin-modified cements

ISO 4049:2009 - Dentistry -- Polymer-based restorative materials

and meets the provisions of Council Directive 93/42/EEC concerning Medical Devices which apply to it, and is manufactured in accordance with the technical documentation.

GC FujiCEM 2 is Class IIa according to rule 5 of annex IX of the Council Directive.

Notified Body: British Standards Institution (n° 0086).

Leuven, ...27/01/2015.....
Date

.....
M. Aydin
Head Quality Assurance and
Regulatory Affairs
On behalf of GC EUROPE N.V.