Vericom Co.,Ltd

48, TOEGYEGONGDAN 1-GIL, CHUNCHEON-SI, GANGWON-DO, 200-944, KOREA

EC Declaration of Conformity

Manufacturer : Vericom Co., Ltd
48, Toegyegongdan 1-Gil, Chuncheon-Si, Gangwon-Do, KOREA

European representative : Koo Kyo-Chan
3 Rue Rosiers 92340 Bourg la Reine France

Product Name : Root Canal filling Material
Brand Name: Well-Root PT


Classification : Class IIa by Rule 8 of Annex IX, MDD 93/42/EEC as amended by Directive 2007/47/EC
Conformity Assessment Route : Annex II, Excluding section 4, Directive 93/42/EEC as amended by Directive 2007/47/EC

Applied harmonized standards, national standards or other normative documents
: EN ISO13485:2016/AC:2016, EN ISO14971:2012, EN ISO10993-1:2009/AC:2010, ISO6876:2012, EN1041:2008, EN1641:2009, EN ISO 15223-1:2016, MEDDEV 2.12-1(rev.8), MEDDEV2.7.1(rev.4)

We herewith declare that the above mentioned products meet the provisions of the Council Directive 93/42/EEC as amended by Directive 2007/47/EC for medical devices. All supporting documentation is retained under the premises of the manufacturer.

Notified Body : SGS Belgium NV, SGS House Noorderlaan 87 2030 Antwerp Belgium
Number of Notified Body: 1639
EC Certificate No. KR19/81826325
Start of CE-marking : September 04, 2018
Expiry date of the certificate: 17 September 2021
Place, Date of issue : Gangwon-Do, Korea, 3 February 2020

Signature :



YUNKI KIM / President on behalf of Vericom Co., Ltd.